America

CERTIFICATE

U8 069850 0004 Rev. 00

Holder of Certificate: **Heidolph Instruments GmbH & Co. KG**
Walpersdorfer Str. 12
91126 Schwabach
GERMANY

Certification Mark:

C US

Product: **Laboratory Equipment**

(Hei-VAP Industrial Series)

The product was voluntarily tested according to the relevant safety requirements noted above. It can be marked with the certification mark above. The mark must not be altered in any way. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC 17067. Certification is based on the TÜV SÜD "Testing and Certification Regulations". TÜV SÜD America Inc. is an OSHA recognized NRTL and a Standards Council of Canada accredited certification body.

Test report no.: 028-713121803-000

Date, 2019-01-04

(Abdul Sabbagh)

TÜV SÜD America Inc. • 10 Centennial Drive • Peabody • MA 01960 • USA

ZERTIFIKAT ◆ CERTIFICATE ◆ 認證證書 ◆ СЕРТИФИКАТ ◆ CERTIFICADO ◆ CERTIFICAT
TÜV SÜD

ZERTIFIKAT ◆ CERTIFICATE ◆ 認證證書 ◆ СЕРТИФИКАТ ◆ CERTIFICADO ◆ CERTIFICAT
TÜV SÜD

America

CERTIFICATE

U8 069850 0004 Rev. 00

Model(s):

Hei-VAP Industrial
(May be followed by additonal suffixes, denoting non-safety relevant options)

Tested according to:

UL 61010-1:2012/R:2016-04
UL 61010-2-010:2015
UL 61010-2-051:2015
CAN/CSA-C22.2 No. 61010-1:2012/U2:2016-04
CAN/CSA-C22.2 No. 61010-2-010:2015
CAN/CSA-C22.2 No. 61010-2-051:2015

Production Facility(ies):

069850

Parameters:

Rated Voltage:	1 x 230 V (US Version) or 3 x 400 V (EU Version)
Rated Frequency:	50-60 Hz
Rated Power:	4600 VA
Protection Class:	I

TÜV SÜD America Inc. • 10 Centennial Drive • Peabody • MA 01960 • USA

TÜV®

A4 / 07.17